

Edition Date: Revision Date : Revised by:

# Product Name: Sample Tube Catalogue No.: T500-4AT This document replaces any previous version Image: Catalogue No.: T500-4AT 1. Product Description: Product Description: > Aml Sample Tube with Internal Thread: Disposable, non-sterile, self-standing tube. Caps sold

separately.

- > Compatible with:
  - T500 Screw Cap Series

## 2. Packaging:

Case: 1 000 units per package

## **3. Product Specifications:**

- > Material:
  - Polypropylene
- Temperature range: -196°C to +121°C
- Tubes can be autoclaved in upright position with caps loosened at +121°C, for up to 30 minutes
- Gamma radiation resistant

# 4. Standards and Conformity:

- ISO 2859-1:Sampling and inspection procedures
- FDA: Resin conforms to FDA 21 CFR 177.1520
- USP: Resin conforms to USP Class VI
- > CONEG / RoHS: Plastics and colorants are in conformity with
  - CONEG / RoHS standards for heavy metals
- > **REACH (SVHC):** Plastic is in conformity to REACH standards
- LATEX: Material is Latex Free
- **BSE / TSE:** Material is BSE / TSE Free

### 5. Quality Assurance:

- > Clear, no presence of contamination in plastic
- Visual attributes
- Dimension and volume measurements
- Closure verification
- Leak proof testing in vacuum at 71.3cm Hg
- Gas phase of Liquid Nitrogen resistance



Edition Date: Revision Date : Revised by: 2008-07-22 2018-09-27 Annette Roy, QA Coordinator

# 6. Traceability:

- **Lot No. Composition:** 8 or 9 digits
- > The lot number can be found in one or all of these locations:
  - 1. On exterior case label
  - 2. On label inserted inside the master case
  - 3. On the inner bag

# 7. Storage Conditions:

- > Store at room temperature in normal warehouse conditions
- Avoid temperature variations and humidity
- Protect from any possible contamination
- Protect from any damage to the packaging

## 8. Recommended Use:

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- > Follow chemical resistance chart recommendations
- > For use in automated equipment, follow the equipment manufacturer's instructions
- > Should be used only in the gas phase of Liquid Nitrogen

### **OWNERSHIP OF MATERIALS**:

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.